





Accreditation number by NERA : 041  
 (original Accreditation number by EPA : 041)

**Environmental Technology Research Center**  
**YUAN ZE UNIVERSITY**  
Analysis Report

<b>Client:</b>	Yuann International Technology Co., Ltd	<b>Client Number:</b>	YZ1151C0588
<b>Address:</b>	No.127, Nanshi Sec. 2, Zhongfeng Rd., Pingzhen City, Taoyuan County, Taiwan (R.O.C.)	<b>Sample Base:</b>	Drinking Water
<b>Plant:</b>	Golden Li H.R Consultant Co., Ltd	<b>Test purpose:</b>	<b>Regular inspection</b>
<b>IndustrialType:</b>	---	<b>Sampling Date:</b>	02/02/2026 16:00
<b>Sampling Site:</b>	Xinxing Nanshan No.4 Hall, 1F water dispenser	<b>Sample Received:</b>	02/03/2026
<b>Sampling Organization:</b>	Yuann International Technology Co., Ltd	<b>Report Date:</b>	02/09/2026
<b>Sampling Method:</b>	---	<b>Report Number:</b>	C1150588-1
<b>Sampling Code:</b>	---		

Sample Report Number		C1150203AJ				Analytical Method	Drinking water standard of MoEnV	Remarks
Sample Name(Number)		Drinking Water						
Analyzed Item	Unit	Analytical Value	Start Time of Incubating	Medium	Original Data			
Coliform Group	CFU/100 mL	<1	02/03/2026 11:30	LES Endo Agar	1/0 , 0	NIEA E230.55B	6	
<b>The END</b>								

**PASS**